



**UNIVERSITY OF CALCUTTA**  
**DEPARTMENT OF APPLIED OPTICS AND PHOTONICS**  
**JD-2, Sector III, Salt Lake, Kolkata - 700106**

**Enquiry / Tender for Quotation**

Enquiry No: AOP/5/ENQ/HOD/16-17

Date 02/12/2016

**Cold Circulating Bath**

Item:> 1.

Bath capacity- 20 Lt.

MOC of Bath Interior- Stainless steel 304 quality

MOC of Bath Interior- Power coated CRC sheet

Temperature Range- Ambient to 5°C

Mode of temperature- By a Digital Controller

Accuracy Control- +/- 0.5°C

Pump Flow Rate- 4 to 6 Lt/minute

Pump Suction Pressure- 300 mbar

Operable Voltage- 220/330 V, 50 Hz, S.P, AC Mains

Dear Sir,

I shall be thankful if you quote your price items as overleaf according to the CU rules and following terms & conditions.

- 1) The quotation/bid must be submitted within **Two (02) days** along with your QUOTATION NUMBER, DATE, SEAL & SIGNED BY appropriate authority.
- 2) Our enquiry. no. & date and also its items no. must be mentioned in your quotation.
- 3) Price for which you are quoting must be mentioned along with all applicable taxes and others incidental charges as well as premiums for insurance against all kinds of risks before executing delivery to our Department if applicable.
- 4) Party must also submit copy of their current valid TRADE LICENCE, VAT REGISTRATION, PAN CARD, For manufacturer or sole distributor, they should also submit their valid copy of certificates/ sole authorized distributor certificates with seal & signed by appropriate authority along with quotation.
- 5) For any other queries related to enquiries/tender, you may also visit CU website or mentioned Department.

Thanking you.

**Rajib Chakraborty**  
**Head of the Department**