

UNIVERSITY OF CALCUTTA

Telephone : 91-33-2461-5445/4711/4712
(Ext. 341/342)
Fax No. : 91-33-2461-4849
Email : microbiologyhead@gmail.com



COLLEGE OF SCIENCE
DEPARTMENT OF MICROBIOLOGY
35 Ballygunge Circular Road, (6th Floor)
Kolkata – 700 019. INDIA.

Ref.

Date :

Tender Notice

Reference No. MICRO/BI/E1

Date: January 25, 2018

Sealed tender bids are hereby invited from authorized suppliers/ agents along with necessary business documents for a refrigerated centrifuge. Specifications of the centrifuge are given below.

Intended agencies are requested to submit their technical and financial bids in two separate envelopes in the specified tender box (named 'Tender for Centrifuge', placed in the office room) in the Department of Microbiology, University of Calcutta; 35, B.C. Road, Kolkata 700019.

Bids should be typed and addressed to the Head, Department of Microbiology, University of Calcutta at the above-mentioned address.

Last date and time of submission of bids: 8 Feb, 2018; 2.00pm

Date of Tender evaluation: 8 Feb, 2018; 2.30pm

Any deviation from tender specifications or tender procedure will result in disqualification.

Technical specification of the instrument:

Sl. no	Product description	Quantity
01	Refrigerated Centrifuge with following specifications: • Maximum speed: 16000 rpm (fixed angle) • Temperature range: 0°C to 40°C with temperature display both in °C and °F • Automatic rotor detection • Safety speed limiter function • Setting of speed in RPM and RCF • Short spin function and pre-cool function • Controlled by microprocessor and with choice of acceleration and deceleration levels • Stainless still internal bowl with optimal height for loading and unloading of sample. • Imbalance detection system with automatic stop function • Automatic locking function of the lid • Brushless motor maintenance free and no carbon deposits	01

Handwritten signature
25.1.2018

DR. ALOK KUMAR SIL
Head
Department of Microbiology
University of Calcutta

02	24x2 ml fixed angle rotor for 2 ml microtubes with maximum speed of 15000 rpm, 21000xg	01
03	8x50 ml fixed angle rotor with maximum speed of 6000 rpm, 4800xg	01
04	4x175 ml swing out rotor with maximum speed of 4500 rpm, 3600xg	01
05	6x96 well plate swing out rotor with maximum speed of 3200 rpm, 1350xg	01, Optional
06	Fixed angle rotor for 6 strips of 8X0.2ml with max speed of 10000 rpm, 11000xg	01, Optional

Terms & Conditions:

1. Intended agencies/ suppliers have to submit all inclusive rates including taxes.
2. Bids should be accompanied with self attested copies of current GSTN, Trade License, PAN, Income tax documents, Registration Certificate, with Statutory Authorities and other credentials.
3. The agents/ suppliers should be liable to provide after sale service
 - For the replacement of the damaged parts/accessories during the first year after installation
 - For the regular maintenance during second and third year after installation
4. The agents/ suppliers should provide documentary evidence of machines that are working in the same or neighboring institutions as a proof of the reliability of the machine

Acceptance of lowest rate is not obligatory and the undersigned reserves the right to accept or reject any or all quotations without assigning any reason thereof.

Alok Kumar Sil
25.1.2018

DR. ALOK KUMAR SIL
Head
Department of Microbiology
University of Calcutta