



**DEPARTMENT OF CHEMICAL TECHNOLOGY
UNIVERSITY COLLEGES OF SCIENCE & TECHNOLOGY
UNIVERSITY OF CALCUTTA, 92, A. P. C. ROAD, KOLKATA 700 009**

Tender Notice

Ref. No. UGC/SAP/ChemTech/Equipment/10/2018-2019

Date: 26-02-2018

Sealed quotation is invited from the authorized vendors within 11th March'2019 for supplying the following laboratory equipment under specification, to be purchased under UGC DRS1 Scheme No. F.5-9/2016/DRS-I(SAP-II), dated April 19, 2016.

Item	Qty
Constant Temperature Bath	01

Specifications:

- Programmable low temperature precision bath
- **Standard Type:**
 - Temperature control range: -30 - 80°C
 - Temperature accuracy: $\pm 0.05^\circ\text{C}$
 - Bath Capacity : 12 L (Horizontal)
- Voltage stabilizer: 5 KVA
- Instrument Warranty: 2 Years
- Post warranty AMC: 3 Years

.....
(Prof. Achintya Saha)
Coordinator, UGC DRS1
Department of Chemical Technology
University of Calcutta